

ABIDEII-ETH__1

	Anatomical	Rest
Manufacturer	Philips	Philips
Model	Achieva	Achieva
Headcoil	32Ch	32Ch
Field Strength	3T	3T
Sequence	3D TFE	EPI
Flip Angle [Deg]	8	90
Inversion Time (TI) [ms]	1150	-
Echo Time (TE) [ms]	Shortest	25
Repetition Time (TR) [ms]	8.4	2000
Bandwidth per Voxel (Readout) [Hz]	188.3	-
Parallel Acquisition	SENSE x 2.3	-
Partial Fourier	No	-
Number of Slices	180	40
Slice Orientation	Axial	Axial
Slice Phase Encoding Direction	-	-
Slice Acquisition Order	-	-
Slice Thickness [mm]	0.9	3
Slice Gap [mm]	0.09	0
Field of View [mm]	230x230	240x240
Acquisition Matrix	256x256	80x77
Slice In-Place Resolution [mm ²]	0.9x0.9	3x3.1
Number of Measurements	-	210
Acquisition Time [min:sec]	5:46	7.06
Fat Suppression	No	Yes
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm ²]	-	-
Averages	-	-
