

ABIDEII-KUL__3

	Anatomical	Rest
Manufacturer	Philips	Philips
Model	Achieva Ds	Achieva Ds
Headcoil	32Ch	32Ch
Field Strength	3T	3T
Sequence	3D TFE	EPI
Flip Angle [Deg]	8	90
Inversion Time (TI) [ms]	900	-
Echo Time (TE) [ms]	4.6	30
Repetition Time (TR) [ms]	9.4	2500
Bandwidth per Voxel (Readout) [Hz]	131	2188
Parallel Acquisition	SENSE x P1.5 S2.5	SENSE x P2
Partial Fourier	-	-
Number of Slices	182	45
Slice Orientation	Axial	Axial
Slice Phase Encoding Direction	-	-
Slice Acquisition Order	-	-
Slice Thickness [mm]	1.2	2.7
Slice Gap [mm]	-	0.4
Field of View [mm]	250x250	200x200
Acquisition Matrix	208x207	80x78
Slice In-Place Resolution [mm ²]	1.2x1.2	2.5x2.56
Number of Measurements	-	162
Acquisition Time [min:sec]	1:43	7:00
Fat Suppression	No	Yes
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm ²]	-	-
Averages	-	-

Anat Note(s): Reconstructed Voxels 0.98x0.98x1.2

Rest Note(s): Reconstructed voxels 1.56x1.56x2.7