

# ABIDEII-OHSU\_\_1

---

	Anatomical	Rest
Manufacturer	Siemens	Siemens
Model	TriTim	TriTim
Headcoil	12Ch	12Ch
Field Strength	3T	3T
Sequence	3D TFL	EPI
Flip Angle [Deg]	10	90
Inversion Time (TI) [ms]	900	-
Echo Time (TE) [ms]	3.58	30
Repetition Time (TR) [ms]	2300	2500
Bandwidth per Voxel (Readout) [Hz]	180	2298
Parallel Acquisition	-	-
Partial Fourier	No	-
Number of Slices	160	36
Slice Orientation	Axial	-
Slice Phase Encoding Direction	-	-
Slice Acquisition Order	-	-
Slice Thickness [mm]	1.1	3.8
Slice Gap [mm]	0.55	0
Field of View [mm]	256x240	240x240
Acquisition Matrix	256x256	64x64
Slice In-Place Resolution [mm <sup>2</sup> ]	1x1	3.8x3.8
Number of Measurements	-	120
Acquisition Time [min:sec]	9:14	5:07
Fat Suppression	No	Yes
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm <sup>2</sup> ]	-	-
Averages	-	-

---