

# ABIDEII-USM\_1

---

	Anatomical	Rest
Manufacturer	Siemens	Siemens
Model	TriTim	TriTim
Headcoil	12Ch	12Ch
Field Strength	3T	3T
Sequence	MPRAGE	EPI
Flip Angle [Deg]	9	90
Inversion Time (TI) [ms]	900	-
Echo Time (TE) [ms]	2.91	28
Repetition Time (TR) [ms]	900	2000
Bandwidth per Voxel (Readout) [Hz]	240	2894
Parallel Acquisition	-	GRAPPA x 2
Partial Fourier	No	-
Number of Slices	160	40
Slice Orientation	Axial	Axial
Slice Phase Encoding Direction	-	-
Slice Acquisition Order	-	-
Slice Thickness [mm]	1.2	3
Slice Gap [mm]	0.6	0.3
Field of View [mm]	256x240	220x220
Acquisition Matrix	256x256	64x64
Slice In-Place Resolution [mm <sup>2</sup> ]	1x1	3.4x3.4
Number of Measurements	-	240
Acquisition Time [min:sec]	9:14	8.06
Fat Suppression	No	Yes
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm <sup>2</sup> ]	-	-
Averages	-	-

---