

Resting_State	PATIENT POSITION		IMAGING PARAMETERS	
	Patient Entry	Head First	Imaging Mode	2D
	Patient Position	Supine	Pulse Sequence	Gradient Echo
	Coil Configuration	32 Ch Head Coil	Imaging Options	EPI, Asset, FMRI
	Plane	AXIAL	SCAN RANGE	
	Series Description	Resting_State	FOV	22.0
	SCAN TIMING		Slice Thickness	3.4
	Flip Angle	75	Slice Spacing	0.0
	TE	30.0	ACQ TIMING	
	TR	2000.0	Freq	64
	Number of Shots	1	Phase	64
	IMAGE ENHANCE		Freq DIR	R/L
	Filter Choice	None	NEX	1.00
	GATING/TRIGGER		Phase FOV	1.00
	Auto Trigger Type	Off	Auto Shim	Auto
	FMRI		Phase Correction	Yes
	Brain Wave Real Time	2	USER CVS	
	Paradigm String	FACE_LOC	User CV0	1.00
	Paradigm UID	1.2.840.113819.3.69136390 610.1372432428.356	MULTI-PHASE	
	Initial State	Control	Slice per Location	295
	PSD Trigger	Internal	Delay after Acquisition	0
	Slice Order	Interleaved	Seperate Series	0
	View Order	Bottom/Up	Delay after Acquisition without AV	0
	# of Repetitions REST	6	Mask Phase	0
	# of Repetitions ACTIVE	6	Mask Pause	0
	# of Dummy Acquisition	5	DIFFUSION	
	SAT		Recon All Images	On
	Tag Type	None	CONTRAST	
	TRICKS		Contrast Yes/No	No
	Pause On/Off	On		
Auto Subtract	0			
Auto SCIC	Off			

Resting_State