

\\USER\ONR\Basic\Space Fortress\Block 3_auto120BR

Scan Time: 4:04

Voxel size: 3.4×3.4×4.0 [mm]

Rel. SNR: 1.00

SIEMENS: ep2d_bold

Routine	1st Signal/Mode	None
Slice group 1	BOLD	
Slices	t-Test	0
Dist. factor	Threshold	4.00
Position	Window	Growing
Orientation	Dynamic t-maps	1
Phase enc. dir.	Starting ignore meas	0
Rotation	Paradigm size	20
Phase oversampling	Meas[1]	Ignore
FoV read	Meas[2]	Ignore
FoV phase	Meas[3]	Baseline
Slice thickness	Meas[4]	Baseline
TR	Meas[5]	Baseline
TE	Meas[6]	Baseline
Averages	Meas[7]	Baseline
Concatenations	Meas[8]	Baseline
Filter	Meas[9]	Baseline
Coil elements	Meas[10]	Baseline
	Meas[11]	Ignore
Contrast	Meas[12]	Ignore
MTC	Meas[13]	Active
Flip angle	Meas[14]	Active
Reconstruction	Meas[15]	Active
Fat suppr.	Meas[16]	Active
Measurements	Meas[17]	Active
Delay in TR	Meas[18]	Active
Multiple series	Meas[19]	Active
	Meas[20]	Active
Resolution	Motion correction	0
Base resolution	Spatial filter	0
Phase resolution		
Phase partial Fourier		
Filter 1	Sequence	
Raw filter	Introduction	0
Interpolation	Averaging mode	Long term
	Bandwidth	2894 [Hz/Px]
PAT mode	Free echo spacing	0
	Echo spacing	0.4 [ms]
Geometry		
Multi-slice mode	EPI factor	64
Series	RF pulse type	Normal
	Gradient mode	Fast
Special sat.		
System		
Scan at current TP		
Scan region position		
Scan region position		
MSMA		
Sagittal		
Coronal		
Transversal		
Head 3T / HE		
Shim mode		
Confirm freq. adjustment		
Assume Silicone		
Ref. amplitude [1H]		
Adjust volume		
Position		
Orientation		
Rotation		
R >> L		
A >> P		
F >> H		
Physio		