

SIEMENS MAGNETOM TrioTim syngo MR B17

\USER\SWU-fMRI\Qiu Jiang\GBB_Multi_mode_1\rest

TA: 8:08 PAT: Off Voxel size: 3.4x3.4x3.0 mm Rel. SNR: 1.00 SIEMENS: ep2d_pace

Properties

| | |
|---|--------|
| Prio Recon | Off |
| Before measurement | |
| After measurement | |
| Load to viewer | On |
| Inline movie | Off |
| Auto store images | On |
| Load to stamp segments | Off |
| Load images to graphic segments | Off |
| Auto open inline display | Off |
| Start measurement without further preparation | On |
| Wait for user to start | On |
| Start measurements | single |

| | |
|------|-----|
| Body | Off |
| NE2 | Off |
| HEP | On |
| HEA | On |
| SP4 | Off |
| SP2 | Off |
| SP8 | Off |
| SP6 | Off |
| SP3 | Off |
| SP1 | Off |
| SP7 | Off |
| SP5 | Off |

Routine

| | |
|--------------------|-------------|
| Slice group 1 | |
| Slices | 32 |
| Dist. factor | 33 % |
| Position | Isocenter |
| Orientation | Transversal |
| Phase enc. dir. | A > P |
| Rotation | 0.00 deg |
| Phase oversampling | 0 % |
| FoV read | 220 mm |
| FoV phase | 100.0 % |
| Slice thickness | 3.0 mm |
| TR | 2000 ms |
| TE | 30 ms |
| Averages | 1 |
| Concatenations | 1 |
| Filter | None |
| Coil elements | HEA;HEP |

| | |
|--------------------------|----------------|
| Positioning mode | FIX |
| Table position | H |
| Table position | 0 mm |
| MSMA | S - C - T |
| Sagittal | R >> L |
| Coronal | P >> A |
| Transversal | H >> F |
| Coil Combine Mode | Sum of Squares |
| Auto Coil Select | Default |
| Shim mode | Standard |
| Adjust with body coil | Off |
| Confirm freq. adjustment | Off |
| Assume Silicone | Off |
| ? Ref. amplitude 1H | 0.000 V |
| Adjustment Tolerance | Auto |
| Adjust volume | |
| Position | Isocenter |
| Orientation | Transversal |
| Rotation | 0.00 deg |
| R >> L | 220 mm |
| A >> P | 220 mm |
| F >> H | 127 mm |

Contrast

| | |
|-----------------|-----------|
| MTC | Off |
| Flip angle | 90 deg |
| Fat suppr. | Fat sat. |
| Averaging mode | Long term |
| Reconstruction | Magnitude |
| Measurements | 242 |
| Delay in TR | 0 ms |
| Multiple series | Off |

Physio

| | |
|-----------------|------|
| 1st Signal/Mode | None |
|-----------------|------|

BOLD

| | |
|-------------------------|----------|
| GLM Statistics | Off |
| Dynamic t-maps | On |
| Starting ignore meas | 0 |
| Ignore after transition | 0 |
| Model transition states | On |
| Temp. highpass filter | On |
| Threshold | 4.00 |
| Paradigm size | 20 |
| Meas[1] | Baseline |
| Meas[2] | Baseline |
| Meas[3] | Baseline |
| Meas[4] | Baseline |
| Meas[5] | Baseline |
| Meas[6] | Baseline |
| Meas[7] | Baseline |
| Meas[8] | Baseline |
| Meas[9] | Baseline |
| Meas[10] | Baseline |
| Meas[11] | Active |
| Meas[12] | Active |
| Meas[13] | Active |
| Meas[14] | Active |
| Meas[15] | Active |
| Meas[16] | Active |
| Meas[17] | Active |

Resolution

| | |
|-----------------------|--------|
| Base resolution | 64 |
| Phase resolution | 100 % |
| Phase partial Fourier | Off |
| Interpolation | Off |
| PAT mode | None |
| Matrix Coil Mode | Triple |
| Distortion Corr. | Off |
| Prescan Normalize | Off |
| Raw filter | On |
| Elliptical filter | Off |
| Hamming | Off |

Geometry

| | |
|------------------|-------------|
| Multi-slice mode | Interleaved |
| Series | Interleaved |
| Special sat. | None |

System

SIEMENS MAGNETOM TrioTim syngo MR B17

| | |
|-------------------|------------|
| Meas[18] | Active |
| Meas[19] | Active |
| Meas[20] | Active |
| Motion correction | Off |
| Spatial filter | Off |
| Sequence | |
| Introduction | Off |
| Bandwidth | 2232 Hz/Px |
| Free echo spacing | Off |
| Echo spacing | 0.51 ms |
| EPI factor | 64 |
| RF pulse type | Normal |
| Gradient mode | Fast* |